

15/2/16

JHJX Machinery Manufacturing Co.Ltd

CALIBRATION CERTIFICATE

Issued By: JHJX Machinery Manufacturing Co.Ltd
No.99 TongZhou Road
AnHui China 243000

Customer:

Address:

City/State/Zip

TOOL DESCRIPTION

| | | | |
|--------------|------------------------|-------|----------|
| Description: | Electric Torque Wrench | Model | JHEW-08S |
| Max torque | 800Nm | | |
| Accuracy | | | +/- 4% |
| Serial#: | | | |

EQUIPMENT USED:

| | |
|-----------------|---------------|
| Transducer | 扭矩传感器/ RTM-MC |
| Serial# | RTM-MC 05010 |
| Unit Of Measure | Nm |
| Accuracy | +/- 4% |

TESTING PROCEDURE

| | |
|------------------------|-------------|
| Procedure | 3WI-7.6-004 |
| Test Ratio | 0.10% |
| Environment condition: | |
| Temp | 80 Deg F |
| Relative Humidity | 72.1% |

Reference documents for the calibration

JJG 707-2003

Calibration Result:

| Setting torque Nm | READING Nm | Average value Nm |
|--------------------------|-------------------|-------------------------|
| 100 | 95, 98, 99 | 97 |
| 200 | 193, 195, 205 | 197 |
| 300 | 303, 310, 293 | 302 |
| 500 | 513, 508, 503 | 508 |
| 700 | 688, 711, 695 | 698 |
| 800 | 801, 822, 819 | 814 |

Page 2 Of 2

Calibrations Are Performed Using Standards Traceable To NIST. Our Calibration System

Complies With ANSI/NCSL Z540-1 And ISO17025. This Information Applies Only To The Equipment Used Above And May Not Be Reproduced Except In Full ,Without Prior Written Consent. There Is No Implied Warranty That Instrument Will Maintain Its Specified Tolerances During The Calibration Interval Due To Possible Drift, Environment Or Other Factors Beyond Our Control.

Work Performed By

李云坤

Reviewed By

胡井湖

DATE 15/2/16